



EUROPEAN UNION
European Structural and Investment Funds
Operational Programme Research,
Development and Education



Univerzita Palackého
v Olomouci

PALACKÝ UNIVERSITY OLOMOUC Křížkovského 511/8, CZ 771 47 OLOMOUC

Ref.No: 032/PV/OVZ/2021

In Olomouc 01. 04. 2021

Subj.: Explanation of the procurement documentation No.1

According to the public supply contract awarded in the above-threshold regime in open procedure in accordance with provisions of Section 56 of Act No. 134/2016 Coll., Public Procurement Act, as amended (hereinafter referred to as the "Act"), entitled „**IMTM – Cryogenic storage system**“, published in the national Tenders Electronic Daily under Reg. No. of the public procurement: **Z2021-009734** and in the Official Journal of the European Union under No.: **2021/S 058-146212**, we inform you of the following explanation of the procurement documentation according to Section 98 (3) of the Act:

Question No.1:

v bodě 2.2. zadávací dokumentace, v technických požadavcích, zadavatel požaduje:

Format: both tubes and racks

Bude zadavatel akceptovat technické řešení, které umožňuje vkládání samostatných zkumavek (bez racků) do osmimístné předchlazené zásuvky?

Answer:

The contracting authority requires in the technical requirements (Article 2.2. of the procurement documentation): *Format: both tubes and racks*

The question: Will contracting authority accept a technical solution that allows insertion of individual tubes (without racks) into octuple pre-cooled tray?

The answer: Yes, octuple pre-cooled tray belongs to the list of accepted racks. Octuple pre-cooled tray is a subset of all racks.

Due to the fact that the procurement documentation is not changed or supplemented, the contracting authority will not extend the deadline for submission of tenders.

Kind regards

Mgr. Petra Vopálková
contact person for public contract matters